

CATALOG D-104

DIVISION 7

JAN 13 1950

1950

U. S. PATENT OFFICE

359/370




1950




BAUSCH & LOMB PHASE CONTRAST ACCESSORIES

359/370

BAUSCH & LOMB PHASE CONTRAST ACCESSORIES



Including Demonstrations of Comparative Results
of Bright Field and B&L Phase Contrast Techniques




← PHASE CONTRAST

← BRIGHT FIELD

THE THEORY OF PHASE CONTRAST MICROSCOPY

In this catalog you will find a rather complete explanation of the Phase Contrast Method. While a clear conception of this principle is unnecessary for the successful use of the phase accessories, yet the inquiring mind seeks at least a working knowledge of this modern technique.

For those who do not care to go deeply into the scientific explanation there is included a highly condensed definition.



PRACTICAL APPLICATIONS OF PHASE CONTRAST

A full use of Phase Contrast Accessories amply justifies their initial cost. Broadly separating themselves into the general classifications of biology, chemistry, mineralogy and counting, applications of this comparatively new technique are particularly noteworthy in the examination of living cells.

In the following pages are given many examples of practical applications, which, in turn, will suggest many, many more.

B
&L



The Principle of Phase Contrast

UNSTAINED, transparent microscope specimens under standard bright field illumination are generally totally lacking in contrast. In order to bring about contrast so that structural detail may be recognized and studied, a tedious process of preparation has been necessary, often resulting in actual physical distortion of the specimen. The phase contrast method of microscopy overcomes this difficulty. It permits the examination of freshly prepared, or living, unstained material.

Resolving Power of the Microscope

In describing phase contrast microscopy, it is necessary to review, briefly, the theory of image formation in the microscope itself. The microscope is essentially a magnifier, and as such, its general function is to make visible detail which is too small for the unaided eye to see. Magnifying power, however, is not the most important characteristic if the observer is to see detail. More important than magnifying power is resolving power, which is dependent upon the numerical aperture of the microscope objective.

Difficulties With Transparent Materials

Granted sufficient resolving power, considerable difficulty is encountered in the use of conventional microscopy for the examination of fine detail in transparent material. This problem commonly occurs in the examination of many living cells, tissue, and small living organisms which are usually transparent and are mounted in transparent media of almost identical optical properties. The solution of this problem, up to now, has been commonly reached by one of three methods. Stained material may be examined by bright field. Unstained materials may be examined by either or both dark field and polarized light. These methods, however, have their limitations. Staining techniques oftentimes distort the specimen; darkfield sometimes exag-

gerates the contrast of the specimen and the background so that the gradations possessed by the specimen structure are not shown in their proper proportions. Surface layers, rather than internal details, are generally revealed. Polarized light is restricted to anisotropic material.

Phase Contrast, The Answer

It might be generally said that the microscopist desires to study material in as near as possible to its natural state. A new method of illumination, incorporating the principles of phase contrast microscopy, first successfully demonstrated by F. Zernike in 1935, overcomes these difficulties, and is at least a large step toward the ultimate goal. Microscope image formation depends on the physical phenomenon of diffraction. When light falls upon an object having fine structure, the characteristic change brought about may be described as a breaking up of the beam into discrete bundles of rays diverging from the object detail over a wide angle. Diffraction thus occurs at the object plane, and the objective lens receives not only the direct light through the specimen but all other diffracted orders of light falling within the angular limits of the objective aperture. If the source is one providing parallel light, all such bundles entering the objective are imaged at its rear focal plane where they may be observed as a series of maxima and minima diffraction spectra images. In this plane, these images act as second-




Image Formation — Phase Contrast

An annular aperture in the diaphragm, placed in the focal plane of the substage condenser, controls the illumination on the object. The aperture is imaged by the condenser and objective at the rear focal plane, or exit pupil, of the objective. In this plane a phase shifting element, or phase plate, is placed.

Light, shown by the solid lines and undeviated by the object structure, in passing through the phase altering pattern, acquires a one quarter wave length of green light advance over that diffracted by the object structure (broken lines) and passing through that region of the phase plate not covered by the altering pattern. The resultant interference effects of the two portions of light form the final image. Altered phase relations in the illuminating rays, induced by otherwise invisible elements in the specimen, are translated into brightness differences by the phase altering plate.

(The eyepiece is not shown in this diagram.)

ary sources, and send out their own cones of light. Since these images have related phases in this plane, their effect in any other plane may be influenced, controlled, and measured through observable effects. Hence, since all of the light emanates from the same original source, the related spectra become visible through their interference effects wherever they overlap.

It is at this point, in the final image plane, that conditions of structure and phase, as they exist in the original specimen, are reproduced. If a change is introduced in the relative phases of the spectra appearing at the rear focal plane of the objective, the intensity in the final image will vary. Evidence of structure detail will be revealed as a consequence. In the rear focal plane of the objective, a phase difference is introduced in the phase contrast method of microscopy. In this manner, it is possible to suppress, intensify, and even reverse the contrast in the final image, depending upon the nature of the phase shift introduced.

Phase Altering Annuli

Phase altering patterns, commonly referred to as "phase plates," are expedients designed to effect a change in an optical path, or phase relation, between that part of the light entering the objective directly through the object structure, and that part which is diffracted by the structure and also enters the objective. Such phase altering patterns are made by the deposition of films of predetermined thicknesses by high vacuum thermal evaporation processes. These films may be applied to separate glass discs mounted in the objective, or applied directly to a lens surface, depending upon the objective design. Patterns, which have been found to be most practical and effective, are those of the annular type, which introduce a phase shift of one-quarter of a wavelength of green light.

An annular aperture diaphragm is necessary for all types of annular phase altering patterns. It is placed at the

In Brief . . .

On these pages is a detailed explanation of the principle of phase contrast. To readers interested in the mechanics of this technique it will prove very illuminating. For others, primarily concerned with the end result, a condensed explanation is more in order.

Phase Contrast Microscopy is, essentially, a special method of controlled illumination, ideally suited for observing thin, transparent objects whose structural details vary only slightly in thickness and refractive index.

These slight variations are not sufficient in themselves to be visible in the bright field microscope. They do, however, affect light passing through the specimen, causing slight irregularities in the wave fronts. Minor changes in the phase relation of the wave fronts result. Phase contrast equipment transforms these phase changes into corresponding variations of brightness, so that the structure is visible under conditions of enhanced contrast.

front focal plane of the substage condenser and, when illuminated, becomes an effective source at infinity with respect to the object plane. Such an annular source illuminates the object with a hollow cone of light, and controls the obliquity of the light traversing the medium.

The shape of the illuminating aperture is important in the formation of diffraction images, since such images actually determine the nature of the final image through their interference effects, and are images not of the specimen itself, but of the illuminating aperture as influenced by the specimen structure.

For the most critical conditions of illumination, light is focused on the substage condenser annular diaphragm, using the principles of Kohler illumination, to fill it uniformly with light. The real image of the diaphragm is then reproduced through the condenser, object, and objective, at the rear focal plane of the objective.


This, then, satisfies the condition that the object may be considered to be illuminated by a number of point sources located at infinity. Also, at the rear focal plane are imaged the diffraction spectra produced by the object structure.

Formation of Phase Difference

The phase altering annulus is placed in the rear focal plane of the objective and its size is such that it coincides with the image of the substage annulus. When the two annuli are exactly concentric and superimposed, direct transmitted light will pass through the annular region and will differ in phase by one-quarter of a wave length from the light diffracted by the object and entering the objective. This light diffracted by the object passes through the remaining aperture area, which is not covered by the annulus.

Since the effect of the shift may be either to accelerate or retard the phase relation of the directly transmitted light, with respect to the diffracted portion, either positive or negative contrast may result. If the film thickness contributing to the phase shift is such that the directly transmitted light passing through the phase altering element is effectively accelerated over the diffracted light by one-quarter of a wavelength of light, the resulting contrast is referred to as "positive" or "dark" contrast, and the phase altering element is said to be "accelerating."

Positive contrast shows regions of greater optical path in the specimen to



PHOTOMICROGRAPHS

With B&L Phase Contrast Accessories

1 SPIROCHETES OF VINCENT'S ANGINA
Motile, living spirochetes in saliva. Original magnification, 970X.

2 BACTERIA CULTURE
Bacillus infection from beer. Original magnification, 970X.


3 MOLD IN TOMATO PASTE
The cytoplasm of mold hyphae when filled with granules appears very dark under phase contrast. Original magnification, 300X.

4 PROSTATE GLAND, HUMAN
This phase contrast photomicrograph shows one alveolus with concretion in the center. Original magnification, 300X.

5-6 TRACHEA, RAT
Showing hyaline cartilage and ciliated epithelia of the mucous membrane. Original magnification, 300X.

7 ONION SKIN
Phase contrast affords a means of examining the nuclei. Original magnification, 210X.

8 BACTERIA, MEGATHERIUM
These living bacteria under phase contrast show detail as plainly as stained specimens under bright field. Original magnification, 970X.



The Principle of Phase Contrast—Cont.

be darker than surrounding areas. This type of contrast has been found to be preferred because of its general similarity to the regular bright field image, with the added benefit of greatly enhanced contrast and graded variations.

"Negative" or "bright" contrast is produced by a retardation of the undiffracted light with respect to the diffracted light, and regions of greater optical path in the specimen will appear bright against a darker background. The image is thus similar to a dark field image. While such an objective could be easily produced, investigation has disclosed no great advantage to the microscopist of such an approach over other available means of observation.

Since most microscopic specimens exhibit irregular structure, the diffraction spectra are not clearly defined, and the direct image of the illuminating source, only, is sharply defined. This direct light, in addition, is much more intense than the diffracted light.

Metallic Absorbing Film

Since the direct light always completely participates in the image formation, its intensity would tend to overbalance that from the weaker, secondary diffraction maxima. The relative distribution of light between these two determines the amount of contrast available. To compensate for this unbalanced distribution of light, a metallic absorbing film is combined with the phase element to reduce the intensity of the direct light. This absorption tends to equalize the intensities of the two interfering beams and thus increases the sensitivity of contrast by permitting the weak diffracted orders to effect a proportionately greater influence. If the direct beam is not sufficiently dampened (inconel deposit on the phase annulus) to equalize its brightness with that of the diffracted beam, the two will interfere, but there will be an excess of light in the direct beam beyond that necessary for the interference process. That direct light will dilute image contrast. If it is excessive, regular bright field microscopy

Catalog D-104, Page 6


results. The diffracted beam would be drowned out. Similarly, if the diffracted beam is out of balance, the image will be devoid of contrast. The latter results if the inconel deposit be made opaque and the diffracted beam, only, proceeds to the image plane. It is, therefore, obvious that by varying the transmission of the annulus, there can be produced, at will, the complete range from no contrast all the way to a pseudo dark field. The best contrast somewhere in between is that giving a gradation of darkness for the various degrees of densities within the specimen. Bright field gives no contrast at all unless the specimen is thick or absorbs strongly, and dark field gives the extreme black and white contrast occluding everything in between.

Transmission Coating Carefully Chosen

The choice of the balancing inconel transmission coating is therefore a delicate one, and it has been carefully studied. Assuming that the specimen is nearly transparent in bright field illumination, the determination of the transmission is contained in our Bausch & Lomb objectives. If, however, the specimen absorbs strongly in certain areas, a different balance would have to be established. But, in that case, the specimen absorption in itself produces contrast. The phase accessories increase contrast in the non-absorbing sections of the specimen, yielding a desired contrast balance.

Theoretically, monochromatic light of the same wavelength used for the original calibration, should be used when working with phase accessories, if the full phase shift effect is to be realized. It is for this reason that green filters are made available with the Bausch & Lomb Phase Contrast Accessories. The one-quarter wavelength phase shift produced by the B&L accessories is calculated in the green portion of the spectrum. Use of other wavelengths will not produce the same degree of phase shift. If the microscopist chooses to use white light, extinction by interference will not be complete for all colors, since the specimen detail does not exert the same phase altering power for all colors.

Page 7, Catalog D-104



PHOTOMICROGRAPHS

With B&L Phase Contrast Accessories

9 PARAMECIUM
Showing contractile vacuole expanded. Original magnification, 430X.

10 SPERMATOZOON, HUMAN
Head, neck and tail are plainly seen. Original magnification, 970X.

11 MITE
Unstained, whole mount of Demodex folliculorum (follicle mite) from dog. Mounted in balsam. Original magnification, 300X.


12 SARCOMA CELL, HUMAN
In tissue culture. Reticular cell sarcoma. Original magnification, 430X.

13 MOUTH EPITHELIUM
Nuclear detail is well shown under phase contrast. Original magnification, 430X.

14 BLOOD VESSEL, RAT
Cross section. The elastic fibre appears white with phase contrast. Original magnification, 300X.

15 ADIPOSE TISSUE, RAT
Fat has been removed by dehydrating and clearing agents, but the protoplasmic envelopes and nuclei remain. Original magnification, 300X.

16 URINE
Sediment from centrifuged specimen. Phase contrast permits differentiating between types of epithelium and casts, and detecting presence of pus cells and bacteria. Original magnification, 300X.



Practical Uses of Phase Contrast

THE MICROSCOPIST has always had difficulty in observing specimens that were colorless and differed little in optical path (refractive index $\times$ thickness) from their surroundings. In an effort to improve this condition the mineralogist and chemist make use of mounting media of lower or higher indices, giving contrast to the preparation with reference to its background.

Bright-field and dark-field illumination are employed, or polarized light in the case of double refracting material. Although these methods greatly improve the over-all contrast between the preparation and the mounting medium, they fail for the most part to produce contrast within the specimen.

The biologist, in studying living cells, is much more limited in his choice of methods. The mounting media must necessarily be body fluids or other isotonic solutions, that will not cause shrinking or dissolution of the cell protoplasm or permit its escape from the cell.

Unfortunately, all liquids having these characteristics are of low index, slightly above water, and so close to the index of most cell constituents that there is little contrast between the specimen and the mounting medium.

The preferred method for living cells has been the use of non-toxic dyes commonly termed "vital dyes." This method of staining is often difficult and time consuming, and the number of structures of protoplasm that will stain by "vital dyes" is very limited.

Where dyes are of little or no value, the biologist has resorted to the out-of-focus image. For example, if a living epithelium cell from the mouth is examined in saliva at slightly above a good focus, structures such as the nucleus and certain cytoplasmic granules of higher index appear bright. Conversely, cell structures that have a lower index appear dark.

Phase microscopy allows us to examine such differences in index with all structures in the cell at exact focus.

With dark contrast phase objectives, objects of greater optical path than their surroundings appear dark, while those of less optical path present a lighter aspect. The effect obtained is as if the cell were stained with a dye coloring each constituent with an intensity proportional to its optical path.

A number of papers have been published on phase contrast microscopy and further research will reveal many new applications. It is being continually demonstrated that it is applicable to the examination of all transparent material showing slight differences in optical path encountered by the bacteriologist, parasitologist, protozoologist, embryologist, histologist, pathologist, botanist, chemist, and mineralogist.

Briefly stated, a few of these applications classified by subject follow:

Biology

1. The applications of phase contrast in biological science are very many, especially with reference to the living cell. Bacteria in body fluids or cultures usually have a higher index than the liquid and appear dark on a lighter background with dark contrast phase objectives. Spores and flagella, difficult to demonstrate even in fixed dyed preparations, can often be observed in living bacteria. Spirochetes of the mouth and other body fluids usually examined in the living state by means of special dark-field condensers, or after fixation with silver stains, are successfully applicable to the phase method.

Nuclei, cell granules, and cell walls of yeast, mold, and fungi can be observed with much greater contrast, the results dependent somewhat on the age and metabolism of the cell.

Paramecium, amoeba and other protozoa can be studied under phase contrast as living cells with detailed visibility of such structures as nuclei, food and water vacuoles, cilia, and flagella. Smears, stretched, teased or macerated

preparations, and tissue cultures can be examined without the use of dyes, or, in some cases, to advantage in combination with vital or supravital dyes, or Feulgen's stain.

2. Phase contrast microscopy is becoming increasingly valuable for the examination of fixed, stained microtome sections for increasing the detail of structures that are difficult to stain or which, dyed, show little differentiation. It is especially valuable for the study of many neurological preparations, such as nerve cells, fibers and nerve endings, which are usually demonstrated by difficult metallic impregnation methods.

3. Microincinerated zoological and botanical specimens have been very difficult to interpret. While the polarizing microscope has been of some value in this field of biology, phase contrast microscopy, with and without polarized light, is an important additional tool.

Chemistry and Mineralogy

1. Thin sections of rock containing minerals of similar refractive indices such as quartz and feldspars are satisfactorily examined.

2. The refractive index of minerals and chemicals by the immersion method can be more accurately determined.

3. Inclusions in crystals or particles on the surface having a slightly lower or higher optical path than that of the crystal can be readily observed. The degree of purity of a specimen can be readily determined.

4. Phase contrast microscopy can be combined with polarized light by using either a polarizer or analyzer. Double refracting minerals or chemicals that are not readily visible at one optical orientation can be reoriented by use of a circular stage. Reorientation appreciably increases their visibility.

For example, a small crystal of sodium nitrate in a 1.587 index liquid, oriented for that index, would be somewhat difficult to see, even with phase contrast objectives. However, if the crystal be rotated 90 degrees on the circular stage for the 1.336 index, the difference in index between the

specimen and the liquid is such as to give very good contrast. Provided the specimen and liquid are close in index for yellow light, and differ greatly in dispersion, "optical dispersion staining" results.

Counting

1. Particles are necessarily often counted in liquids close to the index of the particle, such as white blood cells in 0.1 N hydrochloric acid. The contrast between the cells and liquid is so small that the "above-focus" method previously mentioned is often used to bring out nuclear structure.

With phase contrast the nucleus and cytoplasm are so well differentiated from the background that a differential count can be obtained without resort to the stained smear method.


Other applications are the counting of red blood cells, blood platelets, bacteria, yeast, mold, plankton, and dust particles. The rulings on glass counting chambers are much more visible. They appear bright on a gray background somewhat similar to that obtained with dark-field illumination.

2. With non-phase contrast objectives, particle aggregates often appear as one particle. With phase contrast, aggregates can very often be separated into individual particles, whose boundaries or other differentiating characteristics are much better shown.

Miscellaneous


1. Surface structure of glass and other transparent materials, such as fine polishing marks and non-homogeneities, are more readily observed with phase contrast than by the usual methods. In the case of opaque material or metals, replicas of such surfaces may be first prepared and then examined with phase contrast objectives.


2. Textiles and paper products are applicable to phase contrast microscopy, both from the standpoint of easier identification and for dyeing, finishing and pigment problems. This method is of value for locating voids and fiber inclusions, as well as in the study of resin and other filler penetration.



COMPARATIVE PHOTOMICROGRAPHS

FELDSPAR		STOMACH, RAT	
1 Phase Contrast Cleavage, perthite and slight alteration in thin section. Original magnification, 430 $\times$. Reduced aperture in bright field.	2 Bright Field	5 Phase Contrast Unstained paraffin section of mucous membrane. Original magnification, 300 $\times$. Reduced aperture in bright field.	6 Bright Field
COUNTING		TESTIS, RAT	
3 Phase Contrast Magnesium fluoride particles, indices 1.378 and 1.390 in isopropyl alcohol, Nd 1.3776. Original magnification, 100 $\times$.	4 Bright Field	7 Phase Contrast Section showing spermatogenesis. Formalin fixation. Mounted in balsam. Original magnification, 430 $\times$.	8 Bright Field





Bausch & Lomb Phase Contrast Accessories

Turret-type Abbe Condenser

SIMPLICITY, convenience and efficiency are the outstanding characteristics of the Bausch & Lomb Turret-type Abbe Condenser Phase Contrast Accessories. Users enthusiastically endorse the Bausch & Lomb policy of designing this equipment so that its successful use requires few variables.

Simplified Equipment

The complete equipment consists of: 1) An assembly of condenser, with iris diaphragm, in stationary stage, and rotatable changer containing four diaphragm stops with annular rings, one for each objective, and one free aperture for bright field use; 2) Four objectives with phase altering patterns in the rear focal planes; 3) Auxiliary telescope; 4) Green filter; 5) Shillaber's oil.

Any B&L laboratory microscope having a rack and pinion substage can be employed. No fuss or bother is entailed in mounting the changer.

The Illustrations

Page 10 (At the left)
The Turret-type Condenser Accessories on Bausch & Lomb Model CTA Binocular Microscope.

Page 11 (Below)

Left: The complete accessories in the leatherette-covered case. These can be ordered complete as shown, with the four objectives. In case less than the full complement of objectives is desired the basic items, as a group, and only such objectives as are wanted can be ordered individually.

Right: The complete equipment less immersion oil and case.

Objectives

The Phase Contrast Objectives are fitted with standard Society threads and screw into the nosepiece exactly the same as do the conventional type. Available in 16mm, 8mm, 4mm, and 1.8mm, they give magnifications of 10 $\times$, 21 $\times$, 43 $\times$, and 97 $\times$, (oil immersion).

Telescope

For pre-use centration of the condenser diaphragm image and the phase altering pattern of the objectives an auxiliary telescope is supplied. This adjustment for concentricity of the two rings is quickly effected by means of two knurled thumb screws in the turret-type condenser.

Only Three Operations


Full operating instructions are furnished with each set of accessories.


Briefly, there are but three simple, quickly performed steps: 1) Turn the proper objective into the optical path by rotating the nosepiece; 2) Rotate the turret until the corresponding annular stop is in the optical axis, (There is a window opening in the condenser stage through which numbers indicative of the magnifications of the corresponding objectives appear); 3) Center the condenser diaphragm image and the phase altering pattern by using the auxiliary telescope in place of the eyepiece.

Separate Stands Available

Many users have found that it is necessary to keep a microscope for phase contrast work exclusively. Therefore, two types of stands, monocular and binocular, are listed on page 14. On these stands can be placed the B&L Phase Accessories, giving the purchaser the widest possible choice.

SPECIFICATIONS	
Description	Catalog No.
Rotatable, turret-type changer, with four annular stops and one free opening; iris diaphragm; Abbe 1.25 N.A. condenser; centering telescope; and green filter; one dispensing oil bottle; in case	31-58-78-02
PHASE CONTRAST OBJECTIVES Select one or more	
97 $\times$, 1.8mm, 1.25 N.A., oil immersion	31-10-70-02
43 $\times$, 4mm, 0.65 N.A.	31-10-99-01
21 $\times$, 8mm, 0.50 N.A.	31-10-97-01
10 $\times$, 16mm, 0.25 N.A.	31-10-92-01
For basic items under catalog No. 31-58-78-02 and one of each of the four objectives, 31-10-99-01, 31-10-97-01, 31-10-92-01, 31-10-70-02 order	31-58-78-01





Bausch & Lomb Phase Contrast Accessories

Individual Annular Stop Type L.W.D. Condenser

ANSWERING an insistent demand for a condenser for phase contrast examination of tissue cultures in flasks or chambers, Bausch & Lomb has designed the Individual Annular Stop L.W.D. (Long Working Distance) Condenser Phase Contrast Accessories. Utilizing the same "positive" or "dark contrast" as do the Turret-type condenser accessories, the L.W.D. group follows the proven B&L interpretation of Zernike's theory. And results of the use of these accessories have been very highly satisfactory.

For Research or Laboratory Instruments

The condenser has a long working distance of from 8mm to 10mm in air, allowing examination of cultures under natural conditions. This condenser is regularly supplied with a dovetail slide for fitting to any B&L research type microscope with complete centering substage. For use on laboratory type

The Illustrations

Page 12 (At the left)

The Individual Annular Stop L.W.D. Accessories on Bausch & Lomb Model BA Monocular Microscope.

Page 13 (Below)

Left: Like the Turret-type Condenser Equipment, these Individual Annular Stop Type L.W.D. Condenser items are packed in a sturdy, leatherette-covered case.

Right: The complete group less filter, case and immersion oil. Basic items are grouped, while the objectives and their corresponding annular stops are to be ordered individually.

instruments the slide can be unscrewed from the condenser and replaced by an iris diaphragm.

Objectives

Objectives are the same as are employed in the Turret-type accessories, 16mm, 8mm, 4mm, and 1.8mm—magnifications being 10 $\times$, 21 $\times$, 43 $\times$ and 97 $\times$, respectively.

Annular Stops

Individual annular stops are available which correspond to the various objectives. When stops and objectives are ordered together, the annular rings in the stops are accurately centered to the phase patterns in the objectives,


and after the original adjustment require practically no more centering even with frequent change of objectives and stops.


Simplified Equipment

Complete L.W.D. equipment consists of the condenser with two centering keyheads, centering telescope, green filter, dispensing oil bottle, in leatherette-covered carrying case, plus four objectives and four corresponding annular stops. Only such objectives with stops as may be necessary need be ordered, but the balance of the equipment is required. When advisable to assign a special microscope to phase work, choice of a binocular or monocular stand is explained on the next page.

SPECIFICATIONS

Description	Catalog No.
Long Working Distance Condenser, mounted on plate for research type substage*; centering telescope; and green filter; one dispensing oil bottle; in case *Specify if iris diaphragm is wanted in place of research substage plate.	31-58-88-02
PHASE CONTRAST OBJECTIVES	
97 $\times$, 1.8mm, 1.25 N.A., oil immersion.....	31-10-70-02
43 $\times$, 4mm, 0.65 N.A.....	31-10-99-01
21 $\times$, 8mm, 0.50 N.A.....	31-10-97-01
10 $\times$, 16mm, 0.25 N.A.....	31-10-92-01
INDIVIDUAL ANNULAR STOPS	
Specify corresponding stop for each objective ordered.	
For objective	
97 $\times$, 1.8mm, 1.25 N.A., oil immersion.....	31-58-94
43 $\times$, 4mm, 0.65 N.A.....	31-58-93
21 $\times$, 8mm, 0.50 N.A.....	31-58-92
10 $\times$, 16mm, 0.25 N.A.....	31-58-91





Illuminators, Replacements, Stands

ILLUMINATION for Phase Contrast Microscopy is no more complicated a problem than that presented by bright field work. The main requirement, of course, is that the annular ring be evenly and adequately lighted. Bausch & Lomb offers two illuminators which are well suited to this type of work—the Research and the Spherical, both with 6-volt ribbon filament lamps.

Research Illuminator

For either type of Phase Contrast Accessories this instrument is highly satisfactory—the full aperture of the substage condenser is utilized and there is an abundance of light on the specimen. A set of four neutral and one daylight filters is included.

The green filter which is included as a part of each type of B&L Phase Contrast Accessories is preferable for the reason that the phase shift is calibrated in monochromatic light of that wavelength. Therefore, a realization of the greatest benefits in the phase contrast method depends on the green filter's use. This filter fits the filter holder of the Research Illuminator.

The Illustrations

Upper: Bausch & Lomb Research Illuminator with water cell and filter holder, and four neutral and one ground glass filters. Lower: The Spherical Illuminator. A special green filter should be ordered as the one regularly supplied with the phase contrast accessories will not fit the slot in this illuminator.

Transformers for 115-volt, 60-cycle, A.C., are included but not shown in these illustrations.

Spherical Illuminator

Some users prefer the Spherical Illuminator with spherical reflector, and aspheric focusable condenser, and it does a commendable job in furnishing critical illumination for phase contrast. This Bausch & Lomb instrument can also be used for Koehler illumination of the condenser annulus.

A ground glass, as well as a blue filter, is included and, while the 50mm square green filter regularly supplied with the B&L Phase Contrast Accessories will not fit the filter slot of this Spherical Illuminator, there is one listed in the specifications which should be ordered for this instrument.

Microscope Stands

In order to enable the purchaser to exercise the widest latitude in the choice of his Phase Contrast Microscope, Bausch & Lomb supplies two microscope stands with eyepieces but less objectives and substage condensers.

The binocular stand, CTA, is shown on page 10 carrying the Turret-type Condenser Accessories, while the monocular type, BA, appears on page 12, fitted with the Individual Annular Stop L.W.D. Accessories. However, either stand will accommodate either group of accessories.

SPECIFICATIONS

BA Stand, vertical monocular body tube, 39mm diameter; with 10× Huygenian eyepiece; standard B&L rack and pinion coarse adjustment; B&L patent, lever type, side fine adjustment; rack and pinion substage with full 360° ring mount for carrying B&L phase contrast condensers; revolving, dustproof, click stop, quadruple nosepiece; mirror and fork mount; alcohol- and reagent-resistant black finish; chromium plated exposed metal parts; in hardwood carrying case with lock and key; Catalog No. 31-21-55

CTA Stand, interchangeable, inclined binocular body with parallel eyepiece tubes; paired 10× Huygenian eyepieces; standard B&L rack and pinion coarse adjustment; B&L patent, ball bearing side fine adjustment; rack and pinion substage with full 360° ring mount for carrying B&L phase contrast condensers; revolving, dustproof, click stop, quadruple nosepiece; mirror and fork mount; alcohol- and reagent-resistant black finish; chromium plated exposed metal parts; in hardwood carrying case with lock and key; Catalog No. 31-21-88

SPECIFICATIONS	
Description	Catalog No.
ILLUMINATORS	
Spherical Microscope Illuminator with aspheric condenser; blue and ground glass filters; 108-w., 6-v., ribbon filament lamp; iris diaphragm; cord; plug; and transformer for 115-v., 60-cycle, A. C.	31-33-78-54
Green Filter for use in Spherical Illuminator with phase contrast accessories.	31-34-54
Research Microscope Illuminator with corrected condenser; 108-w., 6-v., ribbon filament prefocused base lamp; combination filter and water cell holder; four neutral and one daylight glass filters, 50mm square; and transformer for 115-v., 60-cycle, A. C.; and Pyrex water cell.	31-33-85-11
PHASE CONTRAST ACCESSORIES REPLACEMENTS	
Wratten Filter, green.	42-47-61
Centering Telescope.	31-50-78
Shillaber's Oil, No. 595, in Dispensing Bottle.	31-50-95

TERMS AND PROCEDURE

CATALOG NUMBERS

When ordering, be sure to give catalog number and name of item.

FINANCIAL STANDING

To avoid delay, a purchaser with whom Bausch & Lomb has not previously transacted business should accompany his first order with commercial references or remittance in cash.

TRANSPORTATION CHARGES

Unless otherwise stated on a B&L quotation, all consumer purchases, except repairs, will be shipped transportation charges prepaid to any point within Continental United States of America. Other-than-consumer purchases, all repairs, and goods for delivery outside the above specified area will be shipped f.o.b. Rochester, N. Y., U. S. A. The shipping charges on products sent on memorandum are not prepaid. There is no charge for packing unless otherwise stated. Products made on special order cannot be forwarded C. O. D.

SHORTAGE OR DAMAGE

To avoid error, Bausch & Lomb exercises the utmost care in checking and packing all shipments. All packing should be carefully examined for small items. If a discrepancy is found, report it to Bausch & Lomb immediately. Report to the transportation company at once any claim for loss, damage or breakage. Carriers are responsible for loss or damage to goods in transit. When shipping container shows damage, do not give a clear receipt, but sign for "in bad order". Save the container and obtain an inspection report from the carrier.

RETURNING GOODS

If it is desired to return goods, first send Bausch & Lomb full information by letter. Goods should not be returned for any reason until approval for their return has been received from Bausch & Lomb. Each item should be plainly tagged with sender's name and address. Identification tags will be sent on request.

ELECTRICAL EQUIPMENTS

Electrical equipments for use on voltages other than those listed are available on special order. Send complete specifications for quotations.

INQUIRIES

When writing, please give the date of your order, the Bausch & Lomb order number which appears on the packing slip and, if possible, the Bausch & Lomb invoice number.

PRICES

Prices are subject to change without notice. Orders are subject to acceptance at Rochester, N. Y., U. S. A., at prices prevailing at time of shipment. Orders for standard items may be cancelled if not wanted because of price advance. Excise taxes or other governmental charges will be added wherever applicable. Shipments to points outside Continental United States of America and to its territories, dependencies and possessions, except Hawaii and Alaska, are not subject to excise taxes. Delivery dates are estimated and cannot be guaranteed.

DESIGN CHANGES

Bausch & Lomb reserves the right to make changes in instrument design in accordance with scientific progress, and to supply apparatus resultingly different from that described or illustrated in its literature.



BAUSCH & LOMB OPTICAL CO.

Rochester 2, New York, U. S. A.

Bausch & Lomb products can be obtained from the following offices as well as from distributors and dealers in the United States and in other principal countries throughout the world:

Branch Offices and Subsidiaries

NEW YORK, NEW YORK.....Bausch & Lomb Optical Co., RCA Building, 30 Rockefeller Plaza
CHICAGO, ILLINOIS.....Bausch & Lomb Optical Co., 5 North Wabash Avenue
SAN FRANCISCO, CALIFORNIA....Bausch & Lomb Optical Co., Flood Bldg.
TORONTO, ONTARIO, CANADA....Bausch & Lomb Optical Co., Ltd., 388 Yonge Street
MONTREAL, QUEBEC, CANADA....Bausch & Lomb Optical Co., Ltd., 1449 St. Alexander Street
LONDON, ENGLAND.....Bausch & Lomb Optical Co., Ltd., Africa House, Kingsway, W. C. 2
RIO DE JANEIRO, BRASIL.....Bausch & Lomb do Brasil Ltd., Rua Assembleia, 104-8° Pavt°
São Paulo, Brasil.....Bausch & Lomb do Brasil Ltd., Rua Libero Badaro, 158-8°
BUENOS AIRES, ARGENTINA.....Bausch & Lomb Argentina, S. A., 740 Avenida R. Saenz Peña
STOCKHOLM, SWEDEN.....Aga-Bausch & Lomb, A. B., Malmkillnadsgatan 39

District Offices

BOSTON

DETROIT

LOS ANGELES

PHILADELPHIA

WASHINGTON

BAUSCH & LOMB CATALOGS

THE Bausch & Lomb Optical Company manufactures many types of optical equipment not listed in this catalog. If you have a special optical problem, it is probable that an instrument of standard manufacture can be adapted to your use at moderate cost. New instruments are being developed continually to handle unusual problems. Below are listed some of the many Bausch & Lomb Products in regular production. Send for literature on lines that interest you.

Microscopes and Accessories

Chemical Microscopes
Dairy Microscopes
Dark Field Optical Systems
Euscope (Exton)
Fluorescence Microscope Accessories
Food Inspection Microscopes
Haemacytometers
Laboratory Microscopes
Metallographic Microscopes
Micro-Projectors
Microscope Accessories
Microscope Illuminators
Microtomes
Mold Count Microscopes
Petrographic Microscopes
Phase Contrast Microscope Accessories
Photomicrographic Cameras and Accessories
Polarizing Microscopes
Research Microscopes
Shop Microscope
Toolmaker's Microscope
Vertical Illuminators
Wide Field Stereoscopic Microscopes

Instruments for Measuring

Optical Properties

Abbe and Dipping Refractometers
Colorimetric Apparatus
Opacimeter
Polariscope
Precision Refractometer
Saccharimeters
Spectrographs
Spectrographic Equipment
Spectrometric Equipment
Spectrophotometers

Instruments for Photogrammetry

Autofocus Rectifier
Metrogon Lenses
Multiplex Equipment

Projection Equipment

Cinephor Projection Lenses
Contour Measuring Projectors
Micro-Projectors
Projection Apparatus (Balopticons and Accessories)
2" x 2" Slide Projector
Super-Cinephor Projection Lenses

Ophthalmic Products

Binocular Ophthalmoscope
Clason Visual Acuity Meter
Compact Acuity Projector
Diagnostic Instruments
Ferree-Rand Perimeter
Greens' Refractor
Keratometer
Occupational Vision Service (Ortho-Rater)
Ophthalmic Hydraulic Chair and Unit
Optical Laboratory Equipment
Orthogon "Wide Vision" Lenses
Orthogon Test Lens Set
Panoptik Bifocal Lenses
Poser Slit Lamp
Ray-Ban Sun Glasses
Safety Eyewear
Spectacle Frames, Mountings, and Cases
Universal Slit Lamp

Miscellaneous

Binoculars
Finger Print Magnifier
Magnifiers and Readers
Micro Tessar Lenses
Optical Glass
Photographic Lenses
Searchlight Reflectors
Special Lenses, Prisms, and Reflectors
Spotting Scopes for Riflemen
Telescopes
Optical Instruments for Crime Detection—Glass Control — Metallography — Metal Working — Ceramic Research—Textile Inspection—Chemical Determination—Food and Drug Manufacture—Paper Makers—Paint and Varnish Makers

Glass to meet the exacting specifications for all types of scientific instruments is made in the Bausch & Lomb Glass Plant, the largest optical glass plant in America.

Orthogon "wide vision" eyeglass lenses, designed by the Bausch & Lomb Scientific Bureau, afford perfect correction from center to edge.

BAUSCH & LOMB OPTICAL CO.

ROCHESTER 2, NEW YORK, U. S. A.

New York—Chicago—San Francisco—Montreal—Toronto—London—Rio de Janeiro—São Paulo—Buenos Aires—Stockholm